

Filename:podstawa.plux

Acquisition author: Administrator

Analysis author:

Company name: Sensofar S.L.


Date:2020-12-10 11:22:32

Analysis Software Version: 1.5.2.0

Acquisition

Topography:	2318 x 1936 px	Operator:	Administrator
Area:	65.82 x 54.97 mm	Date:	2020-12-10
Pixel Size:	28.40 µm/pixel	Time:	10:36:09
Technique:	FringeProjection	Measure type:	TopographyColor
Algorithm:	Graycode Slits	Light:	94.60%
Projection light:	21.00%	Measured:	97.94%
Sample ID text:	-	EMM grid:	2 x 2
EMM alg.:	SmartPro	Acq. time:	00:15
Type:	NonContacting	Manufacturer:	Sensofar
Model:	S Wide	SN:	10-011-2020
Version:	SensoSCAN S wide 1.1	Objective:	TELECENTRIC 0.2X
Calib. date:	2020-12-09	Reference:	-
Calib. time:	16:19:11	Magnification:	0.2430

Visualization



ISO 25178 / Height

Sa	10.700 µm	Sq	12.755 µm
Sku	2.3702	Ssk	0.4557
Smean	0.0000 µm	Sv	31.911 µm
Sp	48.166 µm	Sz	80.077 µm





Filename:podstawa.plux

Acquisition author: Administrator

Analysis author:

Company name: Sensofar S.L.

Date:2020-12-10 11:22:32

Analysis Software Version: 1.5.2.0

Comment

-

