

Filename: SKAN 1.plux

Acquisition author: Administrator

Analysis author:

Company name: Sensofar S.L.


Date: 2020-12-10 12:15:41

Analysis Software Version: 1.5.2.0

Acquisition

Topography:	2323 x 1942 px	Operator:	Administrator
Area:	65.96 x 55.14 mm	Date:	2020-12-10
Pixel Size:	28.40 µm/pixel	Time:	12:11:00
Technique:	FringeProjection	Measure type:	TopographyColor
Algorithm:	Graycode Slits	Light:	52.00%
Projection light:	15.00%	Measured:	95.39%
Sample ID text:	-	EMM grid:	2 x 2
EMM alg.:	SmartPro	Acq. time:	00:15
Type:	NonContacting	Manufacturer:	Sensofar
Model:	S Wide	SN:	10-011-2020
Version:	SensoSCAN S wide 1.1	Objective:	TELECENTRIC 0.2X
Calib. date:	2020-12-09	Reference:	-
Calib. time:	16:19:11	Magnification:	0.2430

Visualization



Filename: SKAN 1.plux

Acquisition author: Administrator


Analysis author:

Company name: Sensofar S.L.

Date: 2020-12-10 12:15:41

Analysis Software Version: 1.5.2.0

Profile



ISO 25178 / Height

Sa	31.140 µm	Sq	40.744 µm
Sku	12.4792	Ssk	-1.0279
Smean	0.0000 µm	Sv	892.68 µm
Sp	718.26 µm	Sz	1610.94 µm

ISO 4287 / Amplitude

Ra	27.209 µm	Rsk	-0.5530
Rku	3.3623	Rt	193.29 µm
Rmean	-0.5500 µm	Rv	119.98 µm
Rp	73.312 µm	Rz	193.29 µm
Rq	34.632 µm		

Comment

-

